




OBSOLETE

Obsolete


SIZE		ISSUE 3-4-39	HE.	TITLE: WALL MODEL DENTAL X-RAY UNIT- COUNTERBALANCE SPRING & TUBE SUPP. - LEFT	
KIND	DOW METAL	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	PAINTED	"			
TEMP.		"		CAT. NO.	PART NO. 12346
PATT.	1299	"			



OBsolete
 SEP 17 1975
 DATE.....
 REFER TO.....
 FOR DETAILS SEE E. C. O. #.....
 ENGINEER.....
 DRAFTSMAN.....

SIZE		ISSUE 3-2-39	HE.	TITLE: WALL MODEL DENTAL X-RAY UNIT- COUNTERBALANCE SUPPORT	
KIND	LOW METAL	CHANGED			
SHAPE		7-11-50	J.V.W.		
FINISH	PAINTED	"			
TEMP.		"			
PATT.	1300	"			
H. G. FISCHER & CO. 2935 WABANSIA AVE. CHICAGO, ILL.				CAT. NO.	PART NO. 12375BF

3HLS. - #3X48 TAP - $\frac{3}{16}$ " DEEP
IN ACCORDANCE TO #10X32
TAPPED HOLE - AS SHOWN



2 REQUIRED

DATE **SEP 17 1975**

REFER TO.....


FOR DETAILS SEE E. C. O. #.....

ENGINEER.....

DRAFTSMAN.....

FINISH:
CHROME #2 & POLISH EXCEPT
WHERE MARKED (DO NOT POLISH)

SIZE	BRONZE	ISSUE 1-29-40	J.W.	TITLE: WALL MODEL DENTAL XRAY UNIT COUNTERBALANCE CONNECTING ROD -RIGHT-		
KIND	CASTING	CHANGED				
SHAPE		"				
FINISH	CHROME #2 SEE NOTE	"				
TEMP.		"				
PATT.	1489	"				
H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.				CAT. NO.	PART NO. 12405 $\frac{1}{2}$	




1 REQUIRED

FINISH:
 CHROME #2 & POLISH EXCEPT
 WHERE MARKED (DO NOT POLISH)

OBSOLETE *obsolete*

SIZE	BRONZE	ISSUE 1-29-40	J.W.	TITLE: WALL MODEL DENTAL X-RAY UNIT COUNTERBALANCE CONNECTING ROD -LEFT	
KIND	CASTING	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	CHROME #2 SEE NOTE	"			
TEMP.		"		CAT. NO.	PART NO.
PATT.	1489	"			12406 1/2 F




2 REQUIRED

OBSOLETE

Obsolete

SIZE	$1 \frac{5}{16}$ " O.D. X .081 W.	ISSUE	3-2-39	J.W.	TITLE: WALL MODEL DENTAL "X-RAY" UNIT COUNTERBALANCE TUBE		
KIND	BRASS	CHANGED					
SHAPE	TUBING	"					
FINISH	PAINTED	"					
TEMP.		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
PATT.		"					
					CAT. NO.	PART NO. 12409A	


17 HLS. #18 DRILL CSK.
FOR #8X32 F. HD. SCREW



1 REQUIRED

OBSOLETE

SIZE	$\frac{1}{8}" \times 2"$	ISSUE	3-20-37 J.W.	TITLE:	FISCHER "SP-60" & "DSP-20" 10' RAIL RAIL PLATE		
KIND	C.R. STEEL	CHANGED					
SHAPE		"					
FINISH	CAD. & LAC.	"					
TEMP.	#26	"					
PATT.		"					
H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.				CAT. NO.	PART NO. 10334A		



$\frac{3}{4}$ " DRILL - IN LINE WITH SLOT
AS SHOWN IN TOP VIEW.

INSERT PCE.# 14487 $\frac{1}{2}$ TUBE

SPIN ENDS OVER & FILE TO CONTOUR OF TUBING

SCREW RODS INTO TOP CAP AS SHOWN.

FASTEN TUBE TO TOP CAP WITH #10X32 SCREWS.

FILL WITH LEAD SO THAT TOTAL WEIGHT = 134 LBS. PLUS 1 LB. MINUS $\frac{1}{2}$ LB.

1 REQUIRED

OBSOLETE

ALUMINUM CASTINGS ON VERTICAL CARRIAGE

SIZE		ISSUE 11-13-46 J.V.W.
KIND		CHANGED
SHAPE		" 12-24-46 J.V.W.
FINISH		"
TEMP.		"
PATT.		"


TITLE:

FISCHER "TV-30-96" RAD. & FLUO. UNIT
COUNTERWEIGHT ASSEMBLY

H. G. FISCHER & CO.
2335 WABANSIA AVE.
CHICAGO, ILL.

CAT.
NO.

PART
NO. 14488GA




OBSOLETE

SCALE $\frac{1}{4}$ " = 1"

FROM 8805

obsolete


SIZE		ISSUE 6-5-40 J.W.	TITLE: FISCHER "DS-60" TABLE LEG (DRILLED)	
KIND		CHANGED		
SHAPE		"	H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"		
TEMP.		"		
PATT.		"		
			CAT. NO.	PART NO. 8805F





#11 GA. (.1196") SHEET COLD ROLLED STEEL

1 REQUIRED

SIZE		ISSUE 8-29-46	J.V.W.	TITLE: FISCHER RADIOGRAPHIC TABLE CENTER SUPPORT CHANNEL	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	PAINTED	"			
TEMP.		"		CAT. NO.	
PATT.		"		PART NO.	8809 $\frac{1}{2}$




6 HLS. 1/4" X 20 TAP




2 REQUIRED

OBSOLETE

SIZE		ISSUE 3-30-34	28.	TITLE: FISCHER "D-S-60" TABLE TUBE RAIL BALL BEARING SUPPORT		
KIND	IRON CASTING	CHANGED				
SHAPE		" "		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH	BLK. LACQUER	" "				
TEMP.		" "				
PATT.	7682-693	" "				
				CAT. NO.	PART NO. 8814F	



SIZE	5" SHEET	ISSUE	3-7-34	HW.	TITLE: U.S. 60 UNIT BASE STOP RELEASE LEVER CLAMP PLATE-BLANK	
KIND	SOFT G.R. STEEL	CHANGED				
SHAPE		" "			OBSOLETE	
FINISH		" "				
TEMP.		" "				
PATT.		" "				
H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.					CAT. NO.	PART NO. 8828




7 H.L.S. #17 DRILL CSK.
FOR #8 X 32 F.H.D. SCR.

2 REQUIRED

OBSELETE

SIZE	#14 GA. (.0781")	ISSUE	2-19-34	28.	TITLE: FISCHER "D-S-GO" UNIT PANEL CHANNEL (DRILLED)		
KIND	SHEET STEEL	CHANGED					
SHAPE		" "			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH	BLK. LACQUER	" "					
TEMP.		" "			CAT. NO.		PART NO. 8833F
PATT.		" "					




TOTAL WEIGHT - 41 LBS.

2 REQUIRED

OBSOLETE

SIZE		ISSUE 9-26-35	2/8	TITLE: FISCHER "DS-60" UNIT COUNTERWEIGHT ASSEM.		
KIND		CHANGED				
SHAPE		" "		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH		" "				
TEMP.		" "				
PATT.		" "				
				CAT. NO.	PART NO. 8838G	

2 REQUIRED



OBSOLETE

SIZE	ISSUE 6-18-47 W.J.	TITLE:
KIND	CHANGED	FISCHER "PRECISION 200" X-RAY UNIT
SHAPE	"	BROAD & FINE TRANSF. SECONDARY TUBE
FINISH	"	
TEMP.	"	
PATT.	"	

H. G. FISCHER & CO.	NO.	PART NO.
2338 WABANSIA AVE.		16901
CHICAGO, ILL.		

BLACK SIDE UP



46 HOLES - #18 DRILL

REMOVE GLOSS FROM PANEL
1" DIAMETER ON NATURAL SIDE.

3 HLS. - $\frac{5}{16}$ " DRILL

6 HLS. - #10X32 TAP

REMOVE GLOSS FROM PANEL
 $\frac{3}{4}$ " DIAMETER ON NATURAL SIDE.




SECTION THRU "A"-A
SCALE ~ FULL SIZE

SCALE ~ HALF SIZE

1 REQUIRED

OBSOLETE

SIZE	$\frac{5}{16}$ SHEET	ISSUE	12-12-40 J.V.W.	TITLE:	U.S.A. SHOCKPROOF DENTAL X-RAY UNIT TRANSF. HEAD TANK PANEL		
KIND	LITHARGE IMP.	CHANGED					
SHAPE	BAKELITE	"					
FINISH		"					
TEMP.		"					
PATT.		"			H. G. FISCHER & CO.	CAT. NO.	PART NO.
					2335 WABANSIA AVE.		13616F
					CHICAGO, ILL.		




OBSOLETE

SCALE ~ HALF SIZE

1 REQUIRED

Obsolete
Duplicate

SIZE	$\frac{1}{16}"$ SHEET	ISSUE	12-16-40	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT TRANSF. HEAD TANK PANEL - GASKET		
KIND	NEOPRENE	CHANGED					
SHAPE		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH		"					
TEMP.		"					
PATT.		"					
					CAT. NO.	PART NO. 13617	



Obsolete


SIZE	3/64" SHEET	ISSUE 1-8-41	J.Y.W
KIND	C.R. STEEL	CHANGED	
SHAPE	C.R. STEEL	..	
FINISH	QHR, #2 POLISH ONE SIDE ONLY	..	
TEMP.		..	
PATT.		..	

TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT
TRANSF. HEAD SCREW SHIELD

H. G. FISCHER & CO.
2335 WABANSIA AVE.
CHICAGO, ILL.

CAT
NO.


PART NO. 13618



1 REQUIRED

OBSOLETE
Obsolete


SIZE	.040" SHEET	ISSUE 6-10-41	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT TUBE HEAD - LEAD PROTECTION SHIELD	
KIND	LEAD	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13619



1 REQUIRED

OBSOLETE
Obsolete


SIZE	1/4" SHEET	ISSUE	7-31-41	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT X-RAY TUBE MOUNTING BOARD		
KIND	BLK. BAKELITE	CHANGED					
SHAPE		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH		"					
TEMP.		"					
PATT.		"					
						CAT. NO.	PART NO. 13624



OBSOLETE

SIZE		ISSUE 2-11-41	J.V.W.
KIND	IRON CASTING	CHANGED	
SHAPE		10-23-42	HE.
FINISH	PAINTED	"	
TEMP.		"	
PATT.	1659	"	

TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT COUNTERBALANCE SPG. & TUBE SUPP. - RIGHT		PART NO. 13639AF
H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.	




3 HLS. - # 25 DRILL C'SUNK.
FOR # 6X32 F. HD. SCR.

2 REQUIRED

OBSOLETE

SIZE	IRON	ISSUE 6-27-41	J.V.W.
KIND	CASTING	CHANGED	
SHAPE		"	
FINISH	BLACK #80	"	
TEMP.		"	
PATT.	1663	"	

TITLE:		U.S.A. SHOCKPROOF DENTAL X-RAY UNIT CONTROL BOX-METER HOUSING LOCK RING	
H. G. FISCHER & CO. 2338 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.	PART NO. 13643AF	




OBSOLETE

obsolete


2 REQUIRED

SIZE	#12	ISSUE	6-27-41	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT CONTROL BOX - METER HOUSING		
KIND	ALUM. CRST	CHANGED					
SHAPE		"					
FINISH	PAINTED	"					
TEMP.		"			H. G. FISCHER & CO. 2338 WABANSIA AVE. CHICAGO, ILL.		
PATT.	1664	"					
					CAT. NO.	PART NO. 13644F	




SIZE	IRON	ISSUE	11-25-40	HE.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT- AUTO TRANS. SUPPORT			
KIND	CASTING	CHANGED						
SHAPE		..						
FINISH	BLACK #80	..						
TEMP.		..						
PATT.	1666	..			<table><tr><td>H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.</td><td>CAT. NO.</td><td>PART NO. 13647AF</td></tr></table>	H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.	PART NO. 13647AF
H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.	PART NO. 13647AF						

H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.	PART NO. 13647AF
---	-------------	---------------------



OBSOLETE

SIZE	IRON	ISSUE 11-25-40	IE.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT - AUTO TRANSF. CLAMP	
KIND	CASTING	CHANGED			
SHAPE		"		H. G. FISCHER & CO. CAT. NO. 13648AF	
FINISH	BLACK #80	"			
TEMP.		"		2335 WABANSIA AVE. CHICAGO, ILL.	
PATT.	1667	"			




CENTER STUDS. FILL WITH LEAD
TO BRING TOTAL WEIGHT UP TO-60LBS.

OBSOLETE

Obsolete

SIZE		ISSUE 8-5-41	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY COUNTERWEIGHT ASSEMBLY	
KIND		CHANGED			
SHAPE		"			
FINISH		"			
TEMP.		"			
				H. G. FISCHER & CO. 2335 WABANSIA AVE.	CAT. NO.
					PART NO. 13665G

4 HLS. - 2 HOLES EACH SIDE
 $\frac{1}{4}$ " DRILL - C'SUNK FOR RIVETING




$3\frac{1}{4}$ " O.D. X #16 GA. (.065") WALL SHELBY SEAMLESS STEEL TUBING

2 REQUIRED

OBSOLETE
Obsolete

SIZE		ISSUE 8-4-41	J.V.W.	TITLE: U.S. ARMY SHOCKPROOF DENTAL X-RAY COUNTERWEIGHT TUBE	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13670




OBSOLETE

Obsolete


1 REQUIRED

SIZE	#214	ISSUE	12-27-40 J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT TRANSF. CLAMP & FIL. TRANSF. SURP. - LEFT	
KIND	ALUM. CASTING	CHANGED	5-25-42 J.V.W.		
SHAPE				H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH					
TEMP.					
PATT.	1690				
				CAT. NO.	PART NO. 13676FA




OBSOLETE

SIZE		ISSUE 5-23-41	HE.	TITLE: ARMY DENTAL X-RAY UNIT - CONTROL DIAL PLATE	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13681




OBSOLETE

SIZE	IRON	ISSUE 4-11-42	J.V.W.	TITLE: FISCHER MOD. "AGA" DENTAL X-RAY UNIT CONTROL BOX METER HOUSING	
KIND	CASTING	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	PAINTED	"			
TEMP.		"		CAT. NO.	PART NO.
PATT.	1885	"			13682F



OBSOLETE

SIZE		ISSUE 6-20-41	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT CONTROL BOX-VOLTAGE CHANGE & TERMINAL BOARD	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13685G




$1\frac{5}{8}$ O.D. X #14 GA. (.083") WALL SHELBY SEAMLESS STEEL TUBING

2 REQUIRED

OBSOLETE

SIZE		ISSUE 1-30-41 J.V.W.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT COUNTERBALANCE TUBE	
KIND		CHANGED		
SHAPE		"	H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	PAINTED	"		
TEMP.		"		
PATT.		"		
			CAT. NO.	PART NO. 13686



3 $\frac{1}{2}$ " O.D. X #13 GA. WALL SHELBY SEAMLESS STEEL TUBING

2 REQUIRED

OBSOLETE
O. Isolate

SIZE		ISSUE 1-22-41	J.V.W.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT UPRIGHT GUIDE TUBE	
KIND		CHANGED			
SHAPE		"			
FINISH	CAD. #23	"			
TEMP.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
PATT.		"			
				CAT. NO.	PART NO. 13748F

DRILLED AFTER ASSEM.
#11 DRILL - BOTTOM SIDE ONLY.

THESE HOLES DRILLED
45° EACH SIDE OF ϕ
BOTTOM SIDE.

2 HOLES DRILLED 115° & 2 HOLES
DRILLED 40° EACH SIDE OF ϕ .

6 HOLES - #9 DRILL C'SUNK.
FOR #10X32 F. HD. SCREW.

$\frac{5}{8}$ " DRILL.
BOTTOM SIDE.

DRILLED AFTER ASSEM.
#11 DRILL - TOP SIDE ONLY.

CHROME #2 POLISH
28" OF THIS END ONLY


OBSOLETE

$2\frac{3}{4}$ " O.D. X #11GA. (.120") WALL - SHELBY SEAMLESS STEEL TUBING

1 REQUIRED

SIZE		ISSUE 1-22-41 J.V.W.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT UPRIGHT TUBE	
KIND		CHANGED		
SHAPE		"	H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	SEE NOTE	"		
TEMP.		"		
PATT.		"		
			CAT. NO.	PART NO. 13749F


Obsolete



OBSOLETE

1 REQUIRED



SIZE		ISSUE 6-15-42 J.V.W.	TITLE:	FISCHER "AGA" S.P. DENTAL X-RAY UNIT UPRIGHT TUBE	
KIND		CHANGED			
SHAPE		"			
FINISH	SEE NOTE	"			
TEMP.		"			
PATT.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.
				PART NO.	13749FA



OBSOLETE

1 REQUIRED

SIZE	80-10-10	ISSUE 6-12-41	J.V.W.	TITLE: U.S.A. SHOCKPROOF DENTAL X-RAY UNIT UPRIGHT TUBE SWIVEL BEARING	
KIND	BRONZE	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.	1668	"		CAT. NO.	PART NO. 13750F




4 H.S. ~ $\frac{1}{8}$ " DRILL-CSK.
TO $\frac{3}{16}$ " FACE ON BOTH SIDES.

27 $\frac{3}{32}$ R.

1 $\frac{1}{8}$ R.

1 $\frac{3}{4}$ "




OBSOLETE
obsolete

RIGHT LEVER SHOWN


RIGHT ~ PATT. # 1783
LEFT ~ PATT. # 1784

SIZE	#80-10-10	ISSUE 6-27-41	HE.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT- WHEEL LOCK LEVER - RIGHT & LEFT	
KIND	BRONZE CASTING	CHANGED			
SHAPE		"			
FINISH		"			
TEMP.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO.
DATE		"			
				PART NO.	13751 - LEFT 13752 - RIGHT



RIVET & SMOOTHEN
TO FACE OF CASTING.

PC. # 13754 - DISC.




PC. # 13751 ~ LEFT
PC. # 13752 ~ RIGHT

4 RIVETS - PC. # 13757


RIGHT LEVER SHOWN

1 OF EACH REQUIRED



OBSOLETE

SIZE		ISSUE 7-1-41	IE	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT- WHEEL LOCK LEVER - RIGHT & LEFT	
KIND		CHANGED			
SHAPE		"			
FINISH	CAD. #23	"			
TEMP.		"			
PATT.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO. 13755G - RIGHT 13756G - LEFT



OBSOLETE


1 REQUIRED

SIZE	.040"	ISSUE	11-27-42	HE.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT- TRANSFORMER SHIELD	
KIND	SHEET LEAD	CHANGED				
SHAPE		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"				
TEMP.		"				
PATT.		"				
					CAT. NO.	PART NO. 13779



SIZE		ISSUE 9-25-42	HE.
KIND	IRON CASTING	CHANGED	
SHAPE		"	
FINISH		"	
TEMP.		"	
PATT.	1953	"	


PART NO. 13782F



1 REQUIRED

OBSOLETE


SIZE		ISSUE 9-25-42	HE.	TITLE: FISCHER "AGA" S.P. DENTAL X-RAY UNIT- TRANSF. CLAMP & FIL. TRANSF. SUPP. ~ LEFT	
KIND	IRON CASTING	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2338 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.	1952	"			
				CAT. NO.	PART NO. 13783F



2 REQUIRED

OBSOLETE

SIZE	1 7/8" DIA.	ISSUE 6-15-42	J.V.W.	TITLE: FISCHER MOD. "AGA" S.P. DENTAL X-RAY UPRIGHT GUIDE ROD	
KIND	C.R. STEEL	CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	CHR. W. #3	"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13789




1 REQUIRED

OBSOLETE

obsolete

2 1/4" O.D. X #11 GA. (.120") WALL SHELBY SEAMLESS STEEL TUBING

SIZE		ISSUE 4-10-41	J.V.W.	TITLE: FISCHER MOBILE DENTAL X-RAY UNIT INNER UPRIGHT TUBE	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	SEE NOTE	"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13801




1 REQUIRED

2 $\frac{1}{4}$ " O.D. X #11 GA. (.120") WALL SHELBY SEAMLESS STEEL TUBING

OBSOLETE

obsolete

SIZE		ISSUE 12-8-41	J.V.W.	TITLE: FISCHER MOBILE DENTAL X-RAY UNIT INNER UPRIGHT TUBE	
KIND		CHANGED			
SHAPE		"			
FINISH	SEE NOTE	"			
TEMP.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
PATT.		"			
				CAT. NO.	PART NO. 13801A



1 REQUIRED

2 3/4" O.D. X #13 GA. (.095") WALL SHELBY SEAMLESS STEEL TUBING



OBSOLETE

obsolete

SIZE		ISSUE 4-10-41	J.V.W.	TITLE: FISCHER MOBILE DENTAL X-RAY UNIT OUTER UPRIGHT TUBE-DRILLED	
KIND		CHANGED			
SHAPE		"			
FINISH	DU-LITE #8	"			
TEMP.		"		H. G. FISCHER & CO. 2936 WABANSIA AVE. CHICAGO, ILL.	
PATT.		"			
				CAT. NO.	PART NO. 13802F

TITLE DENTAL FILTER-DIAPHRAGM		H. G. FISCHER & CO. FRANKLIN PARK, ILLINOIS		PART NO. 13904
PATTERN	FINISH	DATE 10-14-51	DR. BY J.V.W.	CH'D.
TOOLS		SCALE FULL		DIMENSIONS FRACTIONAL DECIMAL ANGULAR UNLESS OTHERWISE SPECIFIED
				REVISION
				DATE
				BY

1





OBsolete

OB30/27E

CATALOG # 3568



DEPT.	AMT.	ITEM NO.	NAME	PART NO. OR MATERIAL
S	2	1	ALUM. DISC	.040" SHEET #1100-H14 ALUMINUM
S	1	2	LEAD DISC	1/6" SHEET LEAD
B.O.	3	3	ADHESIVE	#18933
M			ASSEMBLE	

2



- 1
- 2
- 1
- 3


COAT WITH ADHESIVE —
WHEN TACKY PRESS TOGETHER



BLEND EDGES WITH
FILE. MUST FIT
#12370BF CONE HOLDER.

REMOVE ADHESIVE FROM OUTSIDE SURFACES



CUT OFF HANDLE END OF SCREW
JUST BELOW HOLE FOR HANDLE.



1 REQUIRED

OBSOLETE


SIZE	FROM 14717	ISSUE 11-6-42	HE.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT- RAISING SCREW	
KIND		CHANGED			
SHAPE		"			
FINISH		"			
TEMP.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
PATT.		"			
				CAT. NO.	PART NO. 13934



OBSOLETE

1 REQUIRED

SIZE		ISSUE 11-9-42	HE	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT- RAISING GEAR ASSEMBLY	
KIND		CHANGED			
SHAPE		"			
FINISH		"			
TEMP.		"			
				H. G. FISCHER & CO. 2335 WABANSIA AVE.	CAT. NO.
				PART NO. 13935G	




1 REQUIRED

SIZE	FROM 13942G	ISSUE 11-28-42	HE.	TITLE: ARMY SHOCKPROOF DENTAL X-RAY UNIT- TRANSF. SHIELD BRKT. ASSEM. (FORMED)	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13949GF

OBSOLETE



Cancelled



PCE.# 12621

PCE.# 13993F

PCE.# 13961GA



OBSOLETE

See new tracing
6-14-51

SIZE		ISSUE 8-22-41	J.V.W.	TITLE: U.S. NAVY X-RAY UNIT TRANSE. TANK & BEARING PLATE ASSEM.	
KIND		CHANGED 5-5-43	HE.		
SHAPE		"		H. G. FISCHER & CO. 2338 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 13952 GA

1 1/8"

2 1/8"

60"

SCALE ~ $\frac{1}{4}'' = 1''$

1 REQUIRED

OBSOLETE

obsolete

SIZE		ISSUE 6-3-41	J.Y.W.
KIND	FROM 3286	CHANGED	
SHAPE		"	
FINISH		"	
TEMP.		"	
PATT.		"	

TITLE:
U.S. NAVY X-RAY UNIT - FILAMENT TRANSFER
SECONDARY INSULATING STRIP

H. G. FISCHER & CO.
2335 WABANSIA AVE.
CHICAGO, ILL.

CAT.
NO.

PART
NO. 13954

TITLE
FISCHER "ANN" X-RAY UNIT
TANK PANEL GASKET

H. G. FISCHER & CO.
FRANKLIN PARK, ILLINOIS

PART NO.
13963

REVISION
DATE BY

PATTERN
FINISH

SCALE
HALF SIZE

DATE 1-9-53

DR. BY J.V.W.

CHK'D.


DIMENSIONS

TOOLS
CUT ON SQUARE SHEARS
RADIUS PUNCHES #7202 (BIN 2-4)
PUNCH HLS. THRU DRILL JIG #7181
(BIN 2-4)


MATERIAL
1/16" SHEET NEOPRENE
#2065 - #70 DUROMETER TEST

FRACTIONAL
DECIMAL
ANGULAR
UNLESS OTHERWISE SPECIFIED
± .015
± .005
± 1/2°

OBSOLETE



FOLDER #110



4 HRS. #25 DRILL C'SUNK,
FOR #6X32 F. HD. SCREW.


OBSOLETE

SECTION THRU "A-A"
SCALE ∇ FULL SIZE

obsolete

DEVELOPED LENGTH = 17 1/2"
SCALE ∇ HALF SIZE

SIZE		ISSUE 3-31-41	J.V.W.	TITLE: U.S. NAVY X-RAY UNIT TANK-X-RAY SHIELD		
KIND		CHANGED 2-6-41	J.V.W.			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH		"				
TEMP.		"				
PATT.		"				
				CAT. NO.	PART NO. 13965GA	



8-70 HR.
PCE. NO. 12649GA

5-70 H.R.
PCE. NO. 12650GA

6-70 HR.
PCE. NO. 13957G

MATERIAL
2 1/2" O.D. X 2 1/16" I.D. X OPAKE TUBING.
FROM # 13968 C

SIZE		ISSUE 3-31-41	J.V.W.
KIND	SEE NOTE	CHANGED 6-24-42	J.V.W.
SHAPE		" 2-2-43	HE
FINISH		" 3-30-61	R.C.T.
TEMP.		"	
PATT.		"	


TITLE: U.S. NAVY X-RAY UNIT
X-RAY TUBE - SHIELD - FINISHED

H. G. FISCHER & CO.
2335 WABANSIA AVE.
CHICAGO, ILL.

CAT.
NO.

PART
NO. 13968CF



OBSOLETE
see new
Tracing



1 REQUIRED

OBSOLETE


SIZE	# 12	ISSUE 5-10-41	J.V.W.	TITLE: U.S. NAVY X-RAY UNIT CONTROL HOUSING TERMINAL COVER	
KIND	ALUM. CASTING	CHANGED			
SHAPE		"			
FINISH	PAINTED	"			
TEMP.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
PATT.	1749	"			
				CAT. NO.	PART NO. 13983F



1 REQUIRED

OBsolete

SIZE	#214	ISSUE 5-16-41	HE.	TITLE: U.S. NAVY X-RAY UNIT - CONTROL BOX - AUTO TRANSF. CLAMP - UPPER (RIGHT)	
KIND	ALUM. CASTING	CHANGED			
SHAPE		.. 7-13-42	HE.	H. G. FISCHER & CO. CAT. 2335 WABANSIA AVE. NO. CHICAGO, ILL.	
FINISH	BLACK #80	..			
TEMP.		..		PART NO. 13985 FA	
PATT.	1741	..			



OBSOLETE

SIZE	#214	ISSUE	7-13-42	HE	TITLE: U.S. NAVY X-RAY UNIT - CONTROL BOX - AUTO TRANSF. CLAMP - UPPER (LEFT)	
KIND	ALUM. CASTING	CHANGED				
SHAPE		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH	BLACK #80	"				
TEMP.		"			CAT. NO.	
PATT.	1741	"			PART NO.	13985 $\frac{1}{2}$


TITLE FISCHER MOBILE BEDSIDE UNIT TRANSFORMER COVER LEAD STRIP		H.G. FISCHER & CO. FRANKLIN PARK, ILLINOIS		PART NO. - PART OF 13994G	
PATTERN	FINISH	DATE	DR. BY	CHK'D.	REVISION
		6-5-51	W.J.		
TOOLS DRILL JIG # 7214	MATERIAL .040" SHEET LEAD	DIMENSIONS FRACTIONAL DECIMAL ANGULAR UNLESS OTHERWISE SPECIFIED		DATE BY	
		± .015 ± .005 ± 1/2°			

J. J. Obsolete


SEE NOTE ON MAIN
DRAWING

OBSOLETE

8 HLS. # 42 DRILL - C'SINK
TO 3/16" FACE





2 REQUIRED



OBSOLETE


SIZE	#12	ISSUE	5-5-41	J.V.W.	TITLE: U.S. NAVY X-RAY UNIT SHUTTER FRAME		
KIND	ALUM. CASTING	CHANGED					
SHAPE		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH	PAINTED	"					
TEMP.		"			CAT. NO.		PART NO. 13998F
PATT.	1738	"					



OBSOLETE

1 REQUIRED


SIZE	#12	ISSUE	4-15-41	J.V.W.	TITLE:			
KIND	ALUM. CASTING	CHANGED			U.S. NAVY X-RAY UNIT			
SHAPE		"			SHUTTER FRAME ADAPTER			
FINISH	PAINTED	"						
TEMP.		"						
PATT.	1739	"						
					H. G. FISCHER & CO.	CAT. NO.	PART NO.	
					2338 WABANSIA AVE.			1300
					CHICAGO, ILL.			



1 REQUIRED

OBSOLETE


SIZE	1/4" SHEET	ISSUE	4-17-39 J.W.	TITLE:	MOD. "S" SHORT WAVE 6 METER TRANSF. TERM. STRIP-LOWER (FOREIGN)		
KIND	BLK. BAKELITE	CHANGED					
SHAPE		"			H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.		
FINISH		"					
TEMP.		"					
PATT.		"					
				CAT. NO.		PART NO.	12572



1 REQUIRED

OBSOLETE

SIZE	1/4" SHEET	ISSUE 4-17-39	J.W.	TITLE: MOD. "S" S.W. 6 METER - TRANSF. VOLTAGE CHANGE STRIP (FOREIGN) (ENGRAVED)	
KIND	BLK. BAKELITE	CHANGED			
SHAPE		"			
FINISH		"			
TEMP.		"			
PATT.		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	CAT. NO. 12573



OBSOLETE

SIZE		ISSUE 3-25-40	J.W.	TITLE: FISCHER MOD. "GF" S.P. FLUOROSCOPE TRANSF. CABLE CONN. CORD ASSEMBLY	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"		CAT. NO. PART NO. 12580G	
PATT.		"			

#69 (.0292") DRILL

4-V GROOVES FOR
GASKET APPROX. $\frac{1}{64}$ " DEEP

#73 (.024") DRILL

$\frac{1}{2}$ " X 27 THD.

#19 DRILL

6 HLS. #9 DRILL

4 HLS. #10X32 TAP

SCALE $\sim 1" = \frac{1}{2}$ "

1 REQUIRED

"V" GROOVE FOR
GASKET APPROX. $\frac{1}{64}$ " DEEP

SIZE	FROM 10480 $\frac{3}{2}$	ISSUE 6-29-39	J.W.
KIND	BRASS CASTING	CHANGED	
SHAPE		"	
FINISH	BLACK #80	"	
TEMP.		"	
PATT.	8987-1077	"	


TITLE: FISCHER SYNCHROTIMED BUCKY
CYLINDER HEAD VALVE (.5 TO 25 SEC.)

H. G. FISCHER & CO.
2335 WABANSIA AVE.
CHICAGO, ILL.

CAT.
NO.

PART
NO. 12587

OBsolete




SCREW PLATE IN PLACE - 2 HLS-"E" DRILL
 $\frac{1}{8}$ " DEEP - (IN CENTER OF PLATE) - FOR
 $\frac{1}{4}$ " DIAM. C.R. STEEL PIN - $\frac{1}{2}$ " LONG. - DRIVE IN
 PIN & FILE TO CONTOUR OF TUBE.

SCREW RODS INTO TOP CAP
 AS SHOWN. FASTEN TUBE TO
 TOP CAP WITH #10X32 SCREWS.
 FILL WITH LEAD TO MAKE
 TOTAL WEIGHT OF 82 LBS.

1 REQUIRED

OBSOLETE
obsolete

SIZE		ISSUE 8-14-39 J.W.	TITLE:	FISCHER MOD. "TM-30-96" MOBILE SHOCKPROOF X-RAY UNIT - COUNTERWEIGHT ASSEMBLY	
KIND		CHANGED			
SHAPE		"			
FINISH		"			
TEMP.		"		H. G. FISCHER & CO. CAT. 2335 WABANSIA AVE. NO. CHICAGO, ILL.	
PATT.		"			
			PART NO.	12605G	




1 REQUIRED

OBSOLETE

SIZE		ISSUE 8-8-40	J.V.W.	TITLE: FISCHER MOD. "TM-30-96" MOBILE SHOCKPROOF X-RAY UNIT- COUNTERWEIGHT ASSEMBLY	
KIND		CHANGED 9-12-40	J.V.W.		
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT		"			
				CAT. NO.	PART NO. 12605GC

4 HLS. - #9 DRILL - C'SUNK.
TO $\frac{13}{32}$ " FACE.

4 HLS. - #9 DRILL - C'SUNK.
TO $\frac{13}{32}$ " FACE.



$4 \frac{1}{8}$ " O.D. X #11 GA. (.120") WALL SHELBY SEAMLESS STEEL TUBING

1 REQUIRED

OBSOLETE

SIZE		ISSUE 6-7-39	J.W.	TITLE: FISCHER MOD. "TM-30-96" MOBILE SHOCK PROOF X-RAY UNIT COUNTERWEIGHT TUBE	
KIND		CHANGED			
SHAPE		"		H. G. FISCHER & CO. 2335 WABANSIA AVE. CHICAGO, ILL.	
FINISH		"			
TEMP.		"			
PATT.		"			
				CAT. NO.	PART NO. 12608